Instrumentation Solution

The most commonly problems related to the calibration operation are:

- The maintenance manager must "remember" to check the verification table to see which tools need to be verified and when;
- •The forms must be prepared and completed by hand, with the possibility of transcript errors;
- •The error calculation must be done manually with the possibility of typing errors;
- •The list of instruments to be calibrated must be updated manually.

Our Instrumentation Solution, Smart Point[®] Technology-based, meets this need.

Specifically, <u>all instruments are labelled</u> with a Smart Point[®] memory, which contains all the information forming the instrument record card and which are updated for each maintenance and / or calibration operation.



The calibration and maintenance reports are automatically compiled starting from the information included in the instrument maintenance form and from those concerning the sample instrument.

Our Instrumentation solution guarantees the immediate **availability**, directly on the component, of the data necessary for calibration activities, **simplifies** the compilation of reports and minutes, **eliminates** problems related to transcription errors, all with considerable **time savings** and therefore , costs and a significant increase in service **efficiency**.

Instrumentation Solution

The software supplied, besides enabling interaction with Smart Point[®] memory (read and write), permit the simplification of form management and scheduling of activities, and, in the opening phase, automatically signals the interventions expiring.

| 👿 Calibrator - Str | umenti di mis | ura - Rapporto di tarat <i>RAPPORTO</i> | | ATURA STRU | MENTI | | × | I | | | |
|--|--|--|--|--|--------------------|-----------------------------------|--|---|--|--|--|
| N° rapporto | ate Unità DIAT/OPER/NORD/BN | | Data prossima taratura Esito della taratura Prima della taratura I Si C NO Campione di riferimento Denominazione Campione de Modello | | | | | | | | |
| Classe di precisione (+/- % Fs) Casa costruttrice | REG.DI PRES. RTPB 3A1 1 SPRIANO | Campo di misura MBAR Intervallo verifica A Matricola A 27599 | | Denominazione campione Codice campione Matricola Incertezza del campione (Im) %Fs | 005 | campione | SPRIANO | L | | | |
| Unità di appartenenza Risoluzione (Er) %Fs Valori di taratura Incertezza limite richiesta(Ir) +/- (%F.S.) Punto Pressione Valore misurato Valore misurato Errore applicata salita discesa assoluto E percentuale e (%Fs) 0 0 0 0 0 0 0 0 0 0 0 0 0 0 0 0 0 0 0 | | | Controllo dello ze crescente (25-50 | 15/07/2004 ratura effettuata o; verifica n° 4 punt | escludendo l'ultir | h pressione mo punto del fondo | | | | | |
| 50%F.S. 20 75%F.S. 30 100%F.S 40 | | | | Data emission | Cognom | eeNome (C | Controllato da Cognome e Nome) | | | | |
| | cer N° Scheda Codice Strumento Data Intervento | | | | | | Interventi in Scadenza Taratura Esterna N° Scheda Codice Strumento 260 173/NA 09/07/2004 | | | | |
| | | | | | | | Stampa Scadenze Taratura <u>E</u> sterna | | | | |



www.smartpointsa.com business@smartpointsa.com